


温度自动记录器

TP-1000/TP-2000

+125
+100
+80
+40
+10
0.0
-10
-20
-40
-55



记录时间	温度	单位
2003-07-03 11:23:50	26.18	℃
2003-07-03 11:24:00	26.18	℃
2003-07-03 11:24:10	26.18	℃
2003-07-03 11:24:20	26.18	℃
2003-07-03 11:24:30	26.18	℃
2003-07-03 11:24:40	26.18	℃
2003-07-03 11:24:50	26.18	℃
2003-07-03 11:25:00	26.18	℃
2003-07-03 11:25:10	26.18	℃
2003-07-03 11:25:20	26.18	℃
2003-07-03 11:25:30	26.18	℃
2003-07-03 11:25:40	26.18	℃
2003-07-03 11:25:50	26.18	℃
2003-07-03 11:26:00	26.18	℃
2003-07-03 11:26:10	26.18	℃
2003-07-03 11:26:20	26.18	℃
2003-07-03 11:26:30	26.18	℃
2003-07-03 11:26:40	26.18	℃
2003-07-03 11:26:50	26.18	℃
2003-07-03 11:27:00	26.18	℃

www.a-volt.com

可记录高达 **16000** 个温度记录

测温范围：-55°C ~ +125°C (TP-2000)
-100°C ~ +250°C (TP-1000)

测温精度：±0.5°C

温度分辨率：±0.1°C

内置实时时钟

可设置开始记录时间

可设置温度记录时间间隔

不锈钢防水外壳

RS232接口与计算机连接

易于使用的温度记录软件

可生成EXCEL报表文件



温度记录器即使在-40°C的冰冻环境，也能稳定运行！

北京安伏电子技术有限公司

地址：北京市海淀区上地信息路15号玉景大厦

电话：010-62973717, 62965253

传真：010-62965253-8

邮编：100085

